



Fitting **Guide**



 Prior to fitting the mask, remove the elbow from the mask by pressing the side tabs.



 Connect the elbow to the device air tubing.



 With both lower headgear straps released, hold the mask against the face and pull the headgear over the head.



 Bringing the lower straps below the ears, loop the headgear into the lower hooks on the mask frame.



 Unfasten the Velcro[®], pull the upper headgear straps evenly until they are comfortable and firm, and reattach the Velcro.



 Repeat step 5 with the lower headgear straps.



- Attach the combined elbow assembly and air tubing to the mask by pressing the side tabs and pushing into mask frame.
- Turn on the therapy device.



- Adjust the back strap to fine-tune the fitting if necessary.
- Do not overtighten.
- The back strap can be positioned over or under the hair, with hair tied or untied.

Mirage[™] FX for Her

www.resmedmiragefx.com



Distributed by: ResMed Corp 9001 Spectrum Center Boulevard San Diego CA 92123 USA. [EC]REP ResMed (UK) Ltd 96 Milton Park Abingdon Oxfordshire 0X14 4RY UK. See www.resmed.com for other ResMed locations worldwide. Mirage is a trademark of ResMed Ltd and is registered in U.S. Patent and Trademark Office. Velcro is a trademark of Velcro Industries B.V. © 2011 ResMed Ltd. Specifications may change without notice. 1014643/2 2011-11

